

TYPE APPROVAL CERTIFICATE

This is to certify:**That the Hydrostatic Level Probe**

with type designation(s)

SGE-25.Smart (SG-25Smart), SGE-25S.Smart (SG-25S.Smart), SGE-25 (SG-25), SGE-25S (SG-25S), SGE-25C (SG-25C)

Issued to

**APLISENS S.A.
Warszawa, Poland**

is found to comply with

DNV GL rules for classification – Ships, offshore units, and high speed and light craft**Application :****Product(s) approved by this certificate is/are accepted for installation on all vessels classed by DNV GL.****Location classes:**

Temperature	C
Humidity	B
Vibration	B
EMC	B
Enclosure	D

Issued at **Høvik** on **2017-07-03**This Certificate is valid until **2022-06-30**.DNV GL local station: **Gdansk**Approval Engineer: **Nils Jarem**for **DNV GL**Digitally Signed By: Aleks Karlsen
Location: DNV GL Høvik, Norway
Signing Date: 2017-07-04, on behalf of**Odd Magne Nesvåg
Head of Section**

This Certificate is subject to terms and conditions overleaf. Any significant change in design or construction may render this Certificate invalid. The validity date relates to the Type Approval Certificate and not to the approval of equipment/systems installed.



Product description

The following type designations are included in the certificate:

Hydrostatic level probe SGE-25Smart (SG-25Smart)
 Hydrostatic level probe SGE-25S.Smart (SG-25S.Smart)
 Hydrostatic level probe SGE-25 (SG-25)
 Hydrostatic level probe SGE-25S (SG-25S)
 Hydrostatic level probe SGE-25C (SG-25C)

Approval conditions

The Type Approval covers hardware listed under Product description. When the hardware is used in applications to be classed by DNV GL, documentation for the actual application is to be submitted for approval by the manufacturer of the application system in each case. Reference is made to DNV GL rules for classification of ships Pt.4 Ch.9 Control and monitoring systems

Application/Limitation

Ex-certification is not covered by this certificate. Application in hazardous area to be approved in each case according to the Rules and Ex-Certification/ Special Condition for Safe Use listed in valid Ex-certificate issued by a notified/recognized Certification Body.

The cable installation has to be carried out according to the rules for classification of ships with respect to routing, fixing, protection, penetration of decks / bulkheads and the use of cable pipes. Fixing of cables above sensor shall be provided.

Type Approval documentation

User's Manual	DTR.SG...05
APLISENS Technical Specification	Dated 27.08.08, supplemented 15.02.13
Drawings	SG25-A012-TA; SG25-A021-TA PC28S-S001-01; SG25-S002-TA or PC28-A121-TA
APLISENS Product Catalogue	Edition 2013/14
IEL (The Electrotechnical Institute Gdansk) Test Reports	No. 052/LMC-966/2009, dated 25.03.2009 No. 025/LBS-966/2009, dated 03.04.2009 No. 054/LMC-966/2009, dated 25.03.2009 No. 022/LBS-966/2009, dated 03.04.2009 No. 055/LMC-966/2009, dated 25.03.2009 No. 018/LBS-966/2009, dated 31.03.2009 No. 053/LMC-966/2009, dated 25.03.2009 No. 023/LBS-966/2009, dated 03.04.2009 No. 051/LMC-966/2009, dated 25.03.2009 No. 021/LBS-966/2009, dated 03.04.2009 No. 045/LMC-822/2013 dated 14.05.2013 No. 044/LMC-822/2013 dated 14.05.2013

Type approval renewal assessment report for A-13386, DNV GL Gdansk 2017-05-31.

Tests carried out

Applicable tests according to DNV Standard for Certification No. 2.4, April 2006.

Marking of product

The products to be marked with:

- manufacturer name
- model name
- serial number
- power supply ratings

Job Id: **262.1-006045-3**
Certificate No: **TAA000019R**

Periodical assessment

The scope of the periodical assessment is to verify that the conditions stipulated for the type are complied with, and that no alterations are made to the product design or choice of systems, software versions, components and/or materials.

The main elements of the assessment are:

- Ensure that type approved documentation is available
- Inspection of factory samples, selected at random from the production line (where practicable)
- Review of production and inspection routines, including test records from product sample tests and control routines
- Ensuring that systems, software versions, components and/or materials used comply with type approved documents and/or referenced system, software, component and material specifications
- Review of possible changes in design of systems, software versions, components, materials and/or performance, and make sure that such changes do not affect the type approval given
- Ensuring traceability between manufacturer's product type marking and the type approval certificate

Periodical assessment is to be performed after 2 years and after 3.5 years. A renewal assessment will be performed at renewal of the certificate.

END OF CERTIFICATE